

EC Declaration of Conformity

In Accordance with EN ISO/IEC 17050-1:2010

Manufacturer: Nordson EFD LLC
Address: 40 Catamore Boulevard
East Providence, RI 02914 USA
Type of Equipment: Fluid Dispenser
Product Name: PICO *Pulse* and *Touch* Series
Model & Part Number: PICO *Touch* 7361217
 PICO *Pulse* SD 7361218
 PICO *Pulse* HD 7361283
 PICO *Pulse* HD 7362059

Manufacture Date: _____ Serial Number: _____

We hereby declare that the above listed product(s) conform to the essential requirements of the following Directives:

'WEEE Directive' 2002/96/EC
'RoHS Directive' 2011/65/EU
'EMC Directive' 2004/108/EC
'Low Voltage Directive' 2006/95/EC

The following harmonized standards have been applied:

Emissions Standards:

EN 55011, 2007 Industrial, scientific and medical (ISM) radio-frequency equipment - Electromagnetic disturbance characteristics – Limits and methods of measurement, Class
A. EN61000-3-2:2006/A1/A2(2009) Limits for Harmonic Current Emissions
EN61000-3-3:2008 Limitation of voltage changes, voltage fluctuations and flicker in public low voltage supply systems

Immunity Standards:

EN61000-6-2:2005 Electromagnetic compatibility (EMC) – Part 6-2: Generic standards – Immunity for industrial environments.
EN61000-4-2 Electrostatic Discharge (ESD)
EN61000-4-3 Radiated RF Immunity
EN61000-4-4 Electrical Fast Transient/Burst
EN61000-4-5 Surge Immunity
EN61000-4-6 Conducted RF Immunity
EN61000-4-11 Voltage Dips, Short Interruptions

Safety Standards:

EN61010-1 (PM) Safety requirements for laboratory, test & measurement

Nordson EFD LLC

Wayne Burkett

Wayne Burkett
Plant Manager
Place: East Providence, RI USA

Authorized Representative:
European Technical Manager
EFD International
Unit 14, Apex Business Centre
Boscombe Road
Dunstable, Bedfordshire
LU5 4SB England